

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  17D0449946	<b>(X3) Date Survey Completed</b>  01/06/2023
<b>Name of Provider or Supplier</b>  Community Memorial Healthcare	<b>Street Address, City, State</b>  708 N 18th Street, Marysville, KS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2016</b>	<p><b>SUCCESSFUL PARTICIPATION</b> CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on the review of proficiency testing (PT) records and phone interview with the Laboratory Manager on 1/6/2023 at 9:30 a.m., the laboratory failed to successfully participate in PT from College of American Pathologists (CAP) under the specialties: 1. Routine Chemistry for the regulated analytes: a. 0245 Routine Chemistry b. 0315 pH Blood Gas c. 0335 pCO2 Blood Gas 2. See D2096.</p>
<b>D2096</b>	<p><b>ROUTINE CHEMISTRY</b> CFR(s): 493.841(f)</p>

Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on a desk review of proficiency testing (PT) from the provider CAP for Chemistry, and phone interview, the laboratory failed to achieve an acceptable score of 80% or higher for two of three consecutive testing events for the regulated analytes 0245 Routine Chemistry, pH Blood Gas and pCO<sub>2</sub> Blood Gas Findings: 1. Review of CAP PT scores for 2022 revealed the following performance scores for 0245 Routine Chemistry: a. Second testing event 2022 revealed a score of 66%. b. Third testing event 2022 revealed a score of 66% 2. Review of CAP PT scores for 2022 revealed the following performance scores for 0315 pH Blood Gas: a. Second testing event 2022 revealed a score of 60%. b. Third testing event 2022 revealed a score of 60% 3. Review of CAP PT scores for 2022 revealed the following performance scores for 0335 pCO<sub>2</sub> Blood Gas: a. Second testing event 2022 revealed a score of 40%. b. Third testing event 2022 revealed a score of 60% 4. Phone interview with the lab manager on January 6, 2022 at 9:30 a.m. confirmed, the laboratory failed to achieve an acceptable score of 80% or higher for two of three consecutive testing events for the regulated analytes Routine Chemistry, pH Blood Gas and pCO<sub>2</sub> Blood Gas.