

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  17D0451007	<b>(X3) Date Survey Completed</b>  02/15/2024
<b>Name of Provider or Supplier</b>  El Dorado Clinic	<b>Street Address, City, State</b>  700 W Central, Ste 205, El Dorado, KS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5211</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a review of proficiency testing (PT) records from Wisconsin State Laboratory of Hygiene (WSLH), lack of evaluation documentation, and interview with the technical consultant (TC), the laboratory failed to evaluate proficiency testing results for three of six testing events in 2023. Findings: 1. The surveyor requested evaluation documentation for PT results from WSLH. No evaluation documentation was provided at the time of survey for the following events: a. 2023 Hematology 2 b. 2023 Misc QA POC 2 c. 2023 Chem/Endo/Tx 3 2. Interview with the TC on 2/15/24 at 3 p.m. confirmed, the laboratory failed to evaluate proficiency testing results for three of six testing events in 2023.</p>