

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 17D0451291	(X3) Date Survey Completed 04/23/2025
Name of Provider or Supplier William Newton Hospital	Street Address, City, State 1300 E 5th Ave, Winfield, KS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6047	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(8)(i)</p> <p>(b)(8)(i) Direct observations of routine patient test performance, including patient preparation, if applicable, specimen handling, processing and testing;</p> <p>This STANDARD is not met as evidenced by: Based on review of the CMS 209 personnel form, moderate complexity testing personnel (TP) competencies for 2024, and interview with the technical consultant (TC) #1, the technical consultant failed to include direct observation of routine patient test performance in 24 of 45 moderate complexity TP competency assessments. Findings: 1. Review of the CMS 209 revealed 45 moderate complexity TP. 2. The surveyors requested 2024 TP competencies for review. Review of 24 of 24 competencies for rupture of membrane (ROM) TP revealed the dates TC #1 signed the competencies did not match the dates of assessment. 3. The surveyor asked the TC if he performed direct observation of patient testing on the 24 TP. The TC stated he did not perform direct observation of patient testing on any of the 24 ROM TP. 4. In interview on 4/23/25 at 9:25 a.m., TC #1 confirmed, the technical consultant failed to include direct observation of routine patient test performance in 24 of 45 moderate complexity TP competency assessments.</p>