

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 17D0451961	<b>(X3) Date Survey Completed</b> 07/10/2018
<b>Name of Provider or Supplier</b> Wichita Nephrology Group, Pa	<b>Street Address, City, State</b> 818 N Emporia, Suite 310, Wichita, KS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5209</b>	<p><b>PERSONNEL COMPETENCY ASSESSMENT POLICIES</b> CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on lack of documentation of competency training and assessment for testing personnel 1-10 (refer to laboratory personnel report CMS-209) and confirmed by interview with the technical consultant at 10:25 am on July 10, 2018, the laboratory failed to follow their quality assurance policy to assess 10 of 19 testing personnel's competency for both 2016 and 2017.</p>