

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 17D0453069	(X3) Date Survey Completed 08/07/2020
Name of Provider or Supplier Decatur Health System Inc	Street Address, City, State 810 W Columbia St, Oberlin, KS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on review of proficiency testing records the laboratory failed to successfully participate in proficiency testing from American Proficiency Institute for the analyte: alanine aminotransferase (ALT) for two out of three consecutive proficiency testing events: 2019 event 3 and 2020 event 2 (refer D2096)
D2096	ROUTINE CHEMISTRY CFR(s): 493.841(f) Failure to achieve satisfactory performance for the same analyte or test in two

consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on a desk review for proficiency testing from the provider API for Chemistry, the regulated analyte ALT revealed the following performance scores for PT: 1. Third testing event 2019 revealed a score of 40% for ALT 2. Second testing event 2020 revealed a score of 60% for ALT. The laboratory failed to achieve an acceptable score of 80% or higher for two of three consecutive testing events for the analyte:ALT