

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 17D0648340	(X3) Date Survey Completed 12/29/2022
Name of Provider or Supplier Parsons State Hospital & Training Center	Street Address, City, State 2601 Gabriel Ave, Parsons, KS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2010	<p>TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(2)</p> <p>The laboratory must test samples the same number of times that it routinely tests patient samples.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's American Association of Bioanalysts (AAB) proficiency testing (PT) documentation and interview, the laboratory failed to test proficiency samples the same number of times that it routinely tests patient samples. Findings: 1. Review of the laboratory's 2022 PT documentation found the laboratory was testing PT samples multiple times over a 3 day period for 2022 - Q2 Non-Chemistry-Hematology. Five Hematology samples: Vials 6, 7, 8, 9, and 10 were analyzed for 6 regulated analytes of the Complete Blood Count (CBC): Red Bloodcell Count (RBC), Hemoglobin, Hematocrit, Platelet, White Bloodcell Count (WBC), and WBC diff or Cell I.D as follows: a. Sample 6: CBC performed 3 times 5/16/22, 3 times 5/17/22 and 1 time 5/18/22. b. Sample 7: CBC performed 1 time 5/16/22, 1 time 5/17/22 and 1 time 5/18/22. c. Sample 8: CBC performed 1 time 5/16/22, 1 time 5/17/22 and 1 time 5/18/22.. d. Sample 9: CBC performed 2 times 5/16/22, 2 times 5/17/22 and 1 time 5/18/22. e. Sample 10: CBC performed 2 times 5/16/22, 2 times 5/17/22 and 1 time 5/18/22. 2. No documentation for the reason of the repeated testing was provided at survey. The Technical Consultant (TC) was asked if patient samples were routinely tested by multiple runs and multiple day analysis on the same sample. The TC response was she did not know why samples were run multiple times over multiple days. The TC was also the testing personnel for these samples. 3. Interview with TC on 12/29/22 at 10:35 a.m. confirmed, the laboratory failed to test proficiency samples the same number of times that it routinely tests patient samples.</p>