

|  |  |   |
|--|--|---|
| <b>Statement of Deficiencies</b>   | <b>(X1) Provider/Supplier/CLIA Identification Number</b><br>17D0648374 | <b>(X3) Date Survey Completed</b><br>09/27/2018 |
| <b>Name of Provider or Supplier</b><br>Rush County Memorial Hospital   | <b>Street Address, City, State</b><br>801 Locust, La Crosse, KS        |   |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. |  |   |

| <b>(X4) ID Prefix Tag</b> | <b>Summary Statement of Deficiencies</b>  |
|---------------------------|---|
| <b>D5403</b>              | <p>PROCEDURE MANUAL<br/>CFR(s): 493.1251(b)</p> <p>The procedure manual must include the following when applicable to the test procedure: (1) Requirements for patient preparation; specimen collection, labeling, storage, preservation, transportation, processing, and referral; and criteria for specimen acceptability and rejection as described in 493.1242. (2) Microscopic examination, including the detection of inadequately prepared slides. (3) Step-by-step performance of the procedure, including test calculations and interpretation of results. (4) Preparation of slides, solutions, calibrators, controls, reagents, stains, and other materials used in testing. (5) Calibration and calibration verification procedures. (6) The reportable range for test results for the test system as established or verified in 493.1253. (7) Control procedures. (8) Corrective action to take when calibration or control results fail to meet the laboratory's criteria for acceptability. (9) Limitations in the test methodology, including interfering substances. (10) Reference intervals (normal values). (11) Imminently life-threatening test results, or panic or alert values. (12) Pertinent literature references. (13) The laboratory's system for entering results in the patient record and reporting patient results including, when appropriate, the protocol for reporting imminently life threatening results, or panic, or alert values. (14) Description of the course of action to take if a test system becomes inoperable.</p> <p>This STANDARD is not met as evidenced by:<br/>Based on review of procedures, review of patient test reports, and confirmed by interview with the Laboratory Director at 11:50 am on September 27, 2018, it was determined that the laboratory's procedure for Complete Blood Counts (CBCs) failed to include normal CBC reference ranges.</p> |
| <b>D5441</b>              | <p>CONTROL PROCEDURES<br/>CFR(s): 493.1256(a)(b)(c)(g)</p>  |

(a) For each test system, the laboratory is responsible for having control procedures that monitor the accuracy and precision of the complete analytic process. (b) The laboratory must establish the number, type, and frequency of testing control materials using, if applicable, the performance specifications verified or established by the laboratory as specified in 493.1253(b)(3). (c) The control procedures must-- (c)(1) Detect immediate errors that occur due to test system failure, adverse environmental conditions, and operator performance. (c)(2) Monitor over time the accuracy and precision of test performance that may be influenced by changes in test system performance and environmental conditions, and variance in operator performance. (g) The laboratory must document all control procedures performed.

This STANDARD is not met as evidenced by:

A review of Quality Control (QC) records, patient/specimen test logs, and interview with the general supervisor at 11:50 am on September 27, 2018 confirmed the laboratory failed to perform external QC for the Triage Meter Pro for BNP, CKMB, Myoglobin, Troponin I, and D-Dimers before performing patient testing on 55 patient samples.