

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  17D0677930	<b>(X3) Date Survey Completed</b>  01/18/2019
<b>Name of Provider or Supplier</b>  Epiphany Dermatology Of Kansas, Llc	<b>Street Address, City, State</b>  8505 College Blvd, Overland Park, KS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5217</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of proficiency testing for micrographic cutaneous oncology and interview with the office manager on January 18, 2019 at 12:15PM the laboratory failed to verify accuracy twice annually in 2017, 2018 and to date January 18, 2019.</p>
<b>D6127</b>	<p>TECHNICAL SUPERVISOR RESPONSIBILITIES CFR(s): 493.1451(b)(9)</p> <p>The technical supervisor is responsible for evaluating and documenting the performance of individuals responsible for high complexity testing at least semiannually during the first year the individual tests patient specimens.</p> <p>This STANDARD is not met as evidenced by: Based on review of competencies and interview with the office manager the technical supervisor failed to document and evaluate competency for testing personnel #1 in 2018. Findings: 1. Review of competencies showed no initial, semi-annual or annual competency was completed on testing personnel #1 in 2018. 2. Interview with the office manager on January 18, 2018 at 12:15 PM confirmed the technical supervisor failed to evaluate and document competency for testing personnel #1 in 2018.</p>