

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 17D0695714	(X3) Date Survey Completed 08/07/2020
Name of Provider or Supplier Cheyenne County Hospital	Street Address, City, State 210 W First Street, St Francis, KS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on review of proficiency testing records the laboratory failed to successfully participate in proficiency testing from American Proficiency Institute for the analyte: Free thyroxine for two out of three consecutive proficiency testing events: 2020 event 1 and 2020 event 2 (refer D2096)
D2096	ROUTINE CHEMISTRY CFR(s): 493.841(f) Failure to achieve satisfactory performance for the same analyte or test in two

consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on a desk review for proficiency testing from the provider API for Chemistry, the regulated analyte free thyroxine revealed the following performance scores for PT:

1. First testing event 2020 revealed a score of 60% for Free thyroxine
2. Second testing event 2020 revealed a score of 0% for Free thyroxine

The laboratory failed to achieve an acceptable score of 80% or higher for two of three consecutive testing events for the analyte: Free thyroxine