

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 17D0898265	(X3) Date Survey Completed 03/15/2019
Name of Provider or Supplier Wichita Urology Group West	Street Address, City, State 7570 West 21st Street Suite 1014a, Wichita, KS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6045	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(7)</p> <p>(b) The technical consultant is responsible for-- (b)(7) Identifying training needs and assuring that each individual performing tests receives regular in-service training and education appropriate for the type and complexity of the laboratory services performed;</p> <p>This STANDARD is not met as evidenced by: Based on review of competencies and interview with the technical consultant #2 on March 15, 2019 at 10:00 AM confirmed, the technical consultant failed to evaluate and document the initial competency for testing personnel #4.</p>
D6053	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(9)</p> <p>The technical consultant is responsible for evaluating and documenting the performance of individuals responsible for moderate complexity testing at least semiannually during the first year the individual tests patient specimens.</p> <p>This STANDARD is not met as evidenced by: Based on review of competencies and interview with the technical consultant #2 on March 15, 2019 at 10:00 AM confirmed, the technical consultant failed to evaluate and document competency semiannually during the first year of laboratory testing for testing personnel #2.</p>
D6054	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(9)</p>

The technical consultant is responsible for evaluating and documenting the performance of individuals responsible for moderate complexity testing at least annually, after the first year.

This STANDARD is not met as evidenced by:

Based on review of competencies and interview with the technical consultant #2 on March 15, 2019 at 10:00 AM confirmed, the technical consultant failed to evaluate and document the annual competency for testing personnel #1 for 2017 and failed to evaluate and document the annual competency for testing personnel #3 for 2017 and 2018.