Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 17D0899006	**(X3) Date Survey Completed** 08/27/2019
Name of Provider or Supplier Kansas Surgery & Recovery Center, Llc	**Street Address, City, State** 2770 N Webb Rd, Wichita, KS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D5435	MAINTENANCE AND FUNCTION CHECKS CFR(s): 493.1254(b)(2) For equipment, instruments, or test systems developed in-house, commercially available and modified by the laboratory, or maintenance and function check protocols are not provided by the manufacturer, the laboratory must: (i) Define a function check protocol that ensures equipment, instrument, and test system performance that is necessary for accurate and reliable test results and test result reporting. (ii) Perform and document the function checks, including background or baseline checks, specified in paragraph (b)(2)(i) of this section. Function checks must be within the laboratory's established limits before patient testing is conducted. This STANDARD is not met as evidenced by: Request for records of 2018 and 2019 thermometers, and interview with Technicial Personnel #4 on August 27, 2019 at 3:00 pm reveals that the laboratory failed to define a function check protocol for the thermometers used for temperature monitoring. Testing Personnel #4 stated that she had no records of function checks for the thermometers used in the laboratory for the past two years.