

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 17D0981774	(X3) Date Survey Completed 06/22/2018
Name of Provider or Supplier Dermatology Care Center	Street Address, City, State 15 E 11th Street, Liberal, KS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on lack of documents and interview at 08:30 am on 06/22/2018, the laboratory did not verify the accuracy of the direct wet mount preparations and potassium hydroxide (KOH) preparations examinations performed, twice yearly Findings was as follows:. a. Based upon a review of Proficiency the laboratory failed to have any documentation for KOH . which does not meet the standard for twice a year verifying the accuracy. b. This was confirmed in a interview with the Office Manager 06/22 /2018 at 09:30 am</p>
D5413	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b)</p> <p>The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (1) Water quality. (2) Temperature. (3) Humidity. (4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports.</p> <p>This STANDARD is not met as evidenced by: A review of Temperature and humidity logs and interview with staff revealed the laboratory failed to meet the humidity requirements for the laboratory as the Leica</p>

cryostat Findings were as follows: a. Based upon review of manufacture's operator guide for Leica cryostat the laboratory failed to meet the humidity requirement of 0% to 60% and a temperature of 18 to 35 C. . At the time of the survey 06/22/2018 at 9:30 the laboratory failed to produce any documentation for room temperature or humidity