

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 17D1057124	(X3) Date Survey Completed 08/15/2023
Name of Provider or Supplier Tukhs Dermatology Medical Pavilion	Street Address, City, State 2010 Olathe Blvd Pod C, Kansas City, KS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on the review of proficiency testing (PT) records from American Proficiency Institute (API), the laboratory failed to successfully participate in PT under the specialty Microbiology (subspecialty Mycology) for the analyte: Potassium Hydroxide (KOH) for two out of three consecutive testing events. (Refer to D2046)
D2046	MYCOLOGY CFR(s): 493.827(e) Failure to achieve an overall testing event score of satisfactory performance for two

consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on a review of documents of PT scores from the provider API under the specialty Microbiology (subspecialty Mycology) for the analyte KOH, and interview with technical supervisor # 2 (TC#2) revealed the following: 1. Third Event Microbiology 2022 revealed an overall score of 50% for the analyte KOH. 2. Second Event Microbiology 2023 revealed an overall score of 50% for the analyte KOH. 3. An interview with TC#2 on 8/15/23 at 10:30 am. confirmed, the laboratory failed to achieve an acceptable score of 80% or higher for two of three consecutive testing events in the specialty Microbiology, subspecialty Mycology.