Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	17D1068886	07/29/2019
Name of Provider or Supplier	**Street Address, City, State**	
Laboratory Corporation Of America Holdings	12210 West 87th Street, Lenexa, KS	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5213	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(1) The laboratory must verify the accuracy of any analyte or subspecialty without analytes listed in subpart I of this part that is not evaluated or scored by a CMS-approved proficiency testing program. This STANDARD is not met as evidenced by: Based on a review of proficiency testing (PT) records, PT procedure and staff interview reveals that the laboratory failed to evaluate its proficiency testing results that were not scored by the PT program. Findings: 2019 1st event Hematology /Coagulation PT performance evaluation included three ungraded results. 1. Ungraded samples included: Urine Sediment US-02, Vaginal Wet Preparation (KOH) VKP-01 and Urobilinogen UA-01. 2. Lab investigation and evaluation was performed for UA-02 only. No evidence of lab investigation and evaluation was available for US-02 or VKP-01 at the time of survey. 3. LabCorp proficiency testing procedure, page 3 states "If there is a deficiency or an unacceptable result for any survey specimen (graded, non-graded or educational), then investigation and corrective action is required." 4. Interview with Testing Personnel #2 7/29/19 at 12:05 pm confirmed procedure requires investigation of ungraded results and two of three results were not evaluated or investigated.