

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 17D1069021	(X3) Date Survey Completed 07/29/2019
Name of Provider or Supplier Laboratory Corporation Of America Holdings	Street Address, City, State 11164 Noble Drive, Suite 200, Olathe, KS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5213	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(1)</p> <p>The laboratory must verify the accuracy of any analyte or subspecialty without analytes listed in subpart I of this part that is not evaluated or scored by a CMS-approved proficiency testing program.</p> <p>This STANDARD is not met as evidenced by: Based on a review of proficiency testing (PT) records and staff interview reveals that the laboratory failed to evaluate its proficiency testing results that were not scored by the PT program. Findings: 2019 1st event Hematology/Coagulation PT performance evaluation included three ungraded results. 1. Ungraded samples included: Urine Sediment US-02, Vaginal Wet Preparation (KOH) VKP-01 and Urobilinogen UA-01. 2. No evidence of lab investigation and evaluation was available for all three results at the time of survey. 3. Interview with Testing Personnel #1 at 9:50 am 7/29/19 confirmed ungraded results for all three samples were not evaluated or investigated.</p>