

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  17D1088893	<b>(X3) Date Survey Completed</b>  03/22/2023
<b>Name of Provider or Supplier</b>  Advent Health Lenexa	<b>Street Address, City, State</b>  23401 Prairie Star Parkway, Lenexa, KS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D6063</b>	<p><b>LABORATORY TESTING PERSONNEL</b> CFR(s): 493.1421</p> <p>The laboratory must have a sufficient number of individuals who meet the qualification requirements of 493.1423, to perform the functions specified in 493.1425 for the volume and complexity of tests performed.</p> <p>This CONDITION is not met as evidenced by: Based on review of personnel records and interview, the laboratory failed to have educational credentials required to qualify six of twenty-six testing personnel to perform moderate complexity testing (refer to D6065).</p>
<b>D6065</b>	<p><b>TESTING PERSONNEL QUALIFICATIONS</b> CFR(s): 493.1423(b)(1)(2)(3)(4)(i)</p> <p>(b) Meet one of the following requirements: (b)(1) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located or have earned a doctoral, master's, or bachelor's degree in a chemical, physical, biological or clinical laboratory science, or medical technology from an accredited institution; or (b)(2) Have earned an associate degree in a chemical, physical or biological science or medical laboratory technology from an accredited institution; or (b)(3) Be a high school graduate or equivalent and have successfully completed an official military medical laboratory procedures course of at least 50 weeks duration and have held the military enlisted occupational specialty of Medical Laboratory Specialist (Laboratory Technician); or (b)(4)(i) Have earned a high school diploma or equivalent; and</p> <p>This STANDARD is not met as evidenced by: Based on the review of the CMS 209 form, lack of educational records and interview,</p>

the laboratory failed to have the educational credentials required for six of twenty-six testing personnel (TP) to perform moderate complexity laboratory testing at time of survey. Findings: 1. The CMS 209 form provided at survey listed twenty-six moderate complexity TP. 2. No documentation of educational qualifications for TP #15, TP #18, TP#19, TP#22, TP#25 and TP#26 were available for review at the time of survey. 3. The six TP having no educational qualification documents at the time of survey performed 515 creatinine tests on the EPOC analyzer from 3/23/2021 thru 3/22/2023 in the radiology department. 4. Interview with the general supervisor #4 on 3/22/2022 at 1:40 p.m. confirmed the laboratory failed to have the educational credentials required for six of twenty-six TP to perform moderate complexity laboratory testing.