Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	17D2007412	06/15/2022
Name of Provider or Supplier	**Street Address, City, State**	
Kansas Medical Clinic Pa- Dermatology	3511 Clinton Place,Ste C, Lawrence, KS	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5221	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(d) All proficiency testing evaluation and verification activities must be documented. This STANDARD is not met as evidenced by: Based on the review of the laboratory's 2020, 2021 and to date of survey 2022 AAFP proficiency testing evaluations, and interview, the laboratory failed to document proficiency testing evaluation for three of five events the specialty of microbiology. Findings: 1. Review of the laboratory's AAFP proficiency testing evaluations documentation forms from 9/10/20 to 6/15/22 for microbiology had no documentation of the laboratory director's (LD) signature or designee on: a. AAFP PT 2021-A Clinical Microscopy b. AAFP PT 2021-B Clinical Microscopy c. AAFP PT 2022-A Clinical Microscopy 2. The laboratory manager stated she did not know that acceptable results required documentation of the review. 3. The only documentation on the reports was a highlighting of the results. 3. Interview with the laboratory manager on 6/15/22 at 9:40 a.m. confirmed, the laboratory failed to document proficiency testing evaluation activities for three of five events from 9/10/20 to 6/15 /22 in the specialty of microbiology.