

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 17D2092903	(X3) Date Survey Completed 08/23/2019
Name of Provider or Supplier Kansas City Skin & Cancer Center	Street Address, City, State 9301 W 74th Street, Suite 230, Shawnee Mission, KS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2015	<p>TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(5)(6)</p> <p>(5) The laboratory must document the handling, preparation, processing, examination, and each step in the testing and reporting of results for all proficiency testing samples. The laboratory must maintain a copy of all records, including a copy of the proficiency testing program report forms used by the laboratory to record proficiency testing results including the attestation statement provided by the PT program, signed by the analyst and the laboratory director, documenting that proficiency testing samples were tested in the same manner as patient specimens, for a minimum of two years from the date of the proficiency testing event. (6) PT is required for only the test system, assay, or examination used as the primary method for patient testing during the PT event.</p> <p>This STANDARD is not met as evidenced by: Based on a review of WSLH Proficiency testing (PT) during calendar year 2018 and 2019, the laboratory failed to document the proficiency testing samples were handled in the same manner as patient samples. Findings were: A. A review of proficiency testing events during 2018-2019 revealed that the laboratory failed to handle proficiency testing samples in the same manner as patients. During the review, the surveyor found multiple testing of the proficiency testing prior to the submission date of the proficiency events. The Kansas City Skin & Cancer Center's proficiency testing policy states" the proficiency testing will be treated exactly as patient samples, from accessioning through reporting. Tests will be repeated only patient samples would be repeated under the same conditions." B. The testing events reviewed were: WSLH PT 2018-SPMicro-A1-A and WSLH PT 2019-SPMicro-A1 C. The interview with the Laboratory director at 1100 hours confirmed that routinely patient testing would be performed by one testing personnel prior to resulting.</p>