

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  18D0325122	<b>(X3) Date Survey Completed</b>  08/24/2021
<b>Name of Provider or Supplier</b>  Mcdowell Arh	<b>Street Address, City, State</b>  9879 Route 122, Mc Dowell, KY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2094</b>	<p>ROUTINE CHEMISTRY CFR(s): 493.841(e)</p> <p>(1) For any unsatisfactory analyte or test performance or testing event for reasons other than a failure to participate, the laboratory must undertake appropriate training and employ the technical assistance necessary to correct problems associated with a proficiency testing failure. (2) For any unacceptable analyte or testing event score, remedial action must be taken and documented, and the documentation must be maintained by the laboratory for two years from the date of participation in the proficiency testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review of Chemistry proficiency testing results, the laboratory failed to perform corrective actions on 3 testing events with unacceptable scores. Findings include: 1.) Record review revealed that the laboratory scored 80% on the following analytes and there was no documentation of corrective action: 1st event 2020 LDL, 1st event 2021 Sodium, and 2nd event 2021 Cholesterol and Total Iron Binding (TIBC). 2.) Facility has a proficiency testing policy in place that indicates all unacceptable proficiency testing scores are to be evaluated. 3.) Interview with laboratory supervisor on 08/24/2021 at 10:30 AM indicates that the laboratory failed to ensure corrective action was performed on all unacceptable analytes.</p>