

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  18D0325122	<b>(X3) Date Survey Completed</b>  07/26/2022
<b>Name of Provider or Supplier</b>  Mcdowell Arh	<b>Street Address, City, State</b>  9879 Route 122, Mc Dowell, KY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2016</b>	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on review of CMS database proficiency performance report and American Proficiency Institute performance evaluation for 2021 and 2022 the laboratory failed to attain a satisfactory score of at least 80% in routine chemistry for the analyte of pO2 for 2 out of 3 consecutive testing events resulting in unsatisfactory participation. (See D2096)</p>
<b>D2096</b>	<p>ROUTINE CHEMISTRY CFR(s): 493.841(f)</p>

Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on record review of CMS proficiency performance report (155D) and American Proficiency Institute (API) performance evaluation for routine chemistry the laboratory failed to achieve a satisfactory score of at least 80% for the analyte of pO<sub>2</sub> for 2 out of 3 consecutive testing events resulting in unsuccessful performance. Findings: 1) 155D shows 2021 Event 3 Score as 0% pO<sub>2</sub>, 2022 Event 2 Score as 20% pO<sub>2</sub> 2) API shows 2021 Event 3 Score as 0% pO<sub>2</sub>, 2022 Event 2 Score as 20% pO<sub>2</sub>