

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 18D0325122	(X3) Date Survey Completed 03/27/2023
Name of Provider or Supplier Mcdowell Arh	Street Address, City, State 9879 Route 122, Mc Dowell, KY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The following deficiencies are a result of a desk review of proficiency testing scores obtained from the national database and verified with the proficiency testing company. The facility was found to be out of compliance with the conditions of the CLIA program. The following CONDITION LEVEL DEFICIENCIES were found to be out of compliance: D2016 - SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) D6076 - HIGH COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1441
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on a proficiency testing desk review of the Certification and Survey Provider</p>

	<p>Enhanced Reporting (CASPER-0155), College of American Pathologist (CAP) 2021 records (3rd event) and American Proficiency Institute (API) 2022 records (2nd event), the laboratory failed to successfully participate in a proficiency testing program approved by HHS, for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. The laboratory failed to successfully participate in the specialty of Immunohematology for the Compatibility Testing Analyte. (Refer to D2181).</p>
<p>D2181</p>	<p>COMPATIBILITY TESTING CFR(s): 493.863(e)</p> <p>Failure to achieve an overall testing event score of satisfactory for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on a proficiency testing desk review of the CASPER-0155, CAP 2021 records (3rd event, and API) 2022 records (2nd event), the laboratory failed to achieve satisfactory performance (100%) for two of two testing events in the specialty of Immunohematology for the Compatibility Testing analyte. 1. Review of the Casper -0155 report revealed the following: CAP Immunohematology 2021 3rd event Laboratory received an unsatisfactory score of 0% for the Compatibility Testing analyte. API Immunohematology 2022- 2nd Event Laboratory received an unsatisfactory score of 80% for the Compatibility Testing analyte. 2. A review of proficiency testing records from CAP 2021 and API 2022 proficiency testing records confirmed the above findings.</p>
<p>D6076</p>	<p>LABORATORY DIRECTOR CFR(s): 493.1441</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1443 of this subpart and provides overall management and direction in accordance with 493.1445 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on a proficiency testing desk review of the CASPER-0155, CAP 2021 records (3rd event), and API 2022 records (2nd event), revealed that the laboratory director failed to provide overall management and direction of the laboratory services. Refer to D6089.</p>
<p>D6089</p>	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(4)(i)</p> <p>The laboratory director must ensure the proficiency testing samples are tested as required under subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a proficiency testing desk review of the CASPER-0155, CAP 2021 records</p>

(3rd event, and API, 2022 records (2nd event), revealed the laboratory director failed to ensure successful participation in an HHS approved proficiency testing program. Refer to D2181