

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 18D0692926	(X3) Date Survey Completed 10/08/2018
Name of Provider or Supplier Gilbert Barbee Moore & Mcilvoy Psc	Street Address, City, State 201 Park St, Bowling Green, KY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on desk review of Toxicology proficiency testing results from the American Proficiency Institute on 10/08/2018, the laboratory failed to successfully participate in the Digoxin certified analyte in two consecutive testing events. See D2109 and D2118
D2109	TOXICOLOGY CFR(s): 493.845(a) Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.

	This STANDARD is not met as evidenced by: Based on desk review of Toxicology proficiency testing results from the American Proficiency Institute on 10/08/2018, the laboratory failed to attain a satisfactory score of at least 80 percent on the Digoxin certified analyte. Findings include: The laboratory scored 40 percent in the second testing event of 2018 and scored 60 percent in the third testing event of 2018.
D2118	TOXICOLOGY CFR(s): 493.845(f) Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on desk review of Toxicology proficiency testing results from the American Proficiency Institute on 10/08/2018, the laboratory failed to successfully achieve satisfactory performance for the Digoxin certified analyte in two consecutive testing events. Findings include: The laboratory had an unsatisfactory performance in the second event of 2018 with a score of 40 percent and had an unsatisfactory performance in the third event of 2018 with a score of 60 percent.