

| | | |
|--|--|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 18D0705936 | (X3) Date Survey Completed 05/13/2024 |
| Name of Provider or Supplier Sahetya Medical Group | Street Address, City, State 1573 Ky Hwy 259 N, Brownsville, KY | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|--|
| D0000 | The following deficiencies are a result of a desk review of proficiency testing (PT) scores obtained from the national database and verified with the proficiency testing company. The laboratory was found to be out of compliance with the conditions of the CLIA program. The following CONDITION LEVEL DEFICIENCIES were found to be out of compliance: D2016 - 42 C.F.R. 493.803 Condition: Successful participation [proficiency testing] D6000 - 42 C.F.R. 493.1403 Condition: Laboratories performing moderate complexity testing; laboratory director |
| D2016 | <p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on a PT desk review of the Certification and Survey Provider Enhanced</p> |

| | |
|---------------------|---|
| | <p>Reporting (CASPER)-0155 and Wisconsin State Laboratory of Hygiene (WSLH) PT 2023 records (2nd event) and 2024 records (1st event), the laboratory failed to successfully participate in a PT program approved by HHS, for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. The laboratory failed to successfully participate in the specialty of Hematology for the Platelets analyte. (Refer to D2130)</p> |
| <p>D2130</p> | <p>HEMATOLOGY CFR(s): 493.851(f)</p> <p>Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on a PT desk review of the CASPER-0155 and WSLH PT records 2023 (2nd event) and 2024 (1st event), the laboratory failed to achieve satisfactory performance (80% or greater) for 2 of 3 testing events in the specialty of Hematology for the Platelets analyte. 1. A Review of the CASPER-0155 report revealed the following: Hematology 2023- 2nd Event The Laboratory received an unsatisfactory score of 20% for the Platelets analyte. Hematology 2024- 1st Event The Laboratory received an unsatisfactory score of 60% for the Platelets analyte. 2. A review of PT records (2023 and 2024) from WSLH confirmed the above findings.</p> |
| <p>D6000</p> | <p>MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on a PT desk review of the CASPER-0155 Individual Laboratory Report and WSLH PT records 2023 (2nd event) and 2024 (1st event), the laboratory director failed to provide overall management and direction of the laboratory services. (Refer to D6016)</p> |
| <p>D6016</p> | <p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i)</p> <p>The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part;</p> <p>This STANDARD is not met as evidenced by: Based on a PT desk review of the CASPER-0155 and WSLH PT records 2023 (2nd event) and 2024 (1st event), the laboratory director failed to ensure successful participation in an HHS approved PT program. (Refer to 2130)</p> |