

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 18D0875591	(X3) Date Survey Completed 05/03/2022
Name of Provider or Supplier Paducah Dermatology	Street Address, City, State 3101 Parisa Drive, Paducah, KY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Recertification Survey was initiated on 05/03/2022 and concluded on 05/03/2022. The facility was found not to be in compliance with the laboratory requirements of 42 CFR Part 493.1252(d) with deficiencies cited.
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation, staff interview and record review, it was determined the laboratory failed to ensure Potassium Hydroxide (KOH) reagent currently in use had not exceeded the manufacturer's expiration date. Findings include: Review of the laboratory's procedures, on 05/03/2022 at 11:45 AM, revealed there was no documented evidence the laboratory had a system in place to monitor for expired KOH reagents. Observation during a tour of the laboratory, on 05/03/2022 at 11:10 AM, revealed the KOH reagent currently in use had expired on 03/23/2022. Interview with the Technical Consultant, on 05/03/2022 at 11:45 AM, revealed the laboratory did not have a system in place to ensure that reagents exceeding the expiration date were not in use.</p>