

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  18D1092600	<b>(X3) Date Survey Completed</b>  08/05/2019
<b>Name of Provider or Supplier</b>  Advanced Biomedical DbA United Diagnostic Lab	<b>Street Address, City, State</b>  2754 Veach Road, Owensboro, KY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5415</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(c)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies, as appropriate, must be labeled to indicate the following: (1) Identity and when significant, titer, strength or concentration. (2) Storage requirements. (3) Preparation and expiration dates. (4) Other pertinent information required for proper use.</p> <p>This STANDARD is not met as evidenced by: Based on record review and staff interview on 08/05/2019, the laboratory failed to follow manufacturer's requirements for reconstitution of Beckman Coulter Lyophilized calibration material used on the AU 480 chemistry analyzer for the calibration of creatinine. Findings include: 1. The manufacturer's instructions for storage and stability of reconstituted Beckman Coulter Lyophilized Chemistry Calibrators Level 1 and Level 2 state "Reconstituted calibrator materials are stable for seven days from the date of reconstitution when stored at 2-8 degrees C, except for Total and Direct Bilirubin which are stable for four days." 2. The Technical Supervisor acknowledged in an interview at 11:20 AM on 08/05/2019, calibration material used to calibrate creatinine was reconstituted every fourteen days. 3. The Technical Supervisor acknowledged the laboratory failed to have a system in place to ensure the manufacturer's instructions were followed for the reconstitution and storage of calibration materials.</p>