

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 19D1018671	(X3) Date Survey Completed 05/21/2021
Name of Provider or Supplier Rene' Koppel, Md, A Professional Medical Corp	Street Address, City, State 3640 Houma Boulevard, Metairie, LA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Recertification survey was performed on May 21, 2021 at Rene' Koppel MD-A Professional Medical Corp, CLIA ID # 19D1018671. The laboratory was found in compliance with 42 CFR 493 Requirements for Laboratories; however, standard level deficiencies were cited.
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation by surveyor and interview with personnel, the laboratory failed to ensure supplies did not exceed their expiration dates. Findings: 1. Observation by surveyor during laboratory tour on May 21, 2021 at 9:23 am revealed the following expired items: a) Platinum line SubXero Freezing Medium, Lot 80199, Expiration date: 02/21, Quantity: four (4) bottles 2. In interview on May 21, 2021 at 9:25 am, the Histotech confirmed the identified items were expired.</p>
D6087	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(3)(iii)</p> <p>The laboratory director must ensure that laboratory personnel are performing the test methods as required for accurate and reliable results.</p> <p>This STANDARD is not met as evidenced by:</p>

Based on observation by surveyor and interview with personnel, the Laboratory Director failed to ensure the laboratory personnel performed test methods as required. Refer to D5417.