

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 19D2060138	(X3) Date Survey Completed 09/30/2025
Name of Provider or Supplier Skin Surgery Centre Llc	Street Address, City, State 2581 Florida Street, Suite C, Mandeville, LA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Recertification survey was performed at Skin Surgery Centre, LLC, CLIA ID 19D2060138, on September 30, 2025. The laboratory was found in compliance with 42 CFR 493 Requirements for Laboratories; however, standard level deficiencies were cited.
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's policies and interview with personnel, the laboratory failed to establish a complete written competency assessment policy for testing personnel. Findings: 1. Review of the laboratory's "Training and Information" and "Proficiency Testing Competency and CLIA competency assessment" policies revealed the laboratory did not include the following: a) written procedures related to initial training for doctors who serve as testing personnel b) documentation utilized for semi-annual competency assessment for new testing personnel 2. In interview on September 30, 2025 at 11:12 am, the Operations Manager confirmed the laboratory's written competency assessment policy did not include the identified information.</p>
D6103	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(13)</p> <p>(e)(13) Ensure that policies and procedures are established for monitoring individuals who conduct preanalytical, analytical, and postanalytical phases of testing to assure that they are competent and maintain their competency to process specimens, perform test procedures and report test results promptly and proficiently, and whenever</p>

necessary, identify needs for remedial training or continuing education to improve skills;

This STANDARD is not met as evidenced by:

Based on record review and interview with personnel, the Laboratory Director failed to ensure complete policies and procedures for assessing personnel competency were established. Refer to D5209.