

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 21D0014230	(X3) Date Survey Completed 11/30/2018
Name of Provider or Supplier University Of Md Health Center Laboratory	Street Address, City, State Campus Drive Bldg 140 #1163, College Park, MD	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6054	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(9)</p> <p>The technical consultant is responsible for evaluating and documenting the performance of individuals responsible for moderate complexity testing at least annually, after the first year.</p> <p>This STANDARD is not met as evidenced by: Based on review of the annual evaluations and interview with the testing personnel, the technical consultant did not ensure that all testing personnel received an annual competency review by the qualified technical consultant. Findings: 1. The laboratory currently has two testing persons listed on the "Laboratory Personnel Report (CLIA) (CMS-209)." 2. The testing personnel evaluation records for 2018 were reviewed. One of the two testing persons had their annual review performed and documented by a testing person with an associate's degree. The technical consultant/evaluator is required to have a Bachelor of Science degree with two years' experience. 3. During the survey on 11/30/2018 at 12:00 PM the laboratory testing person confirmed that one of the two the annual evaluations for 2018 was not performed by the qualified technical consultant.</p>