

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  21D0210275	<b>(X3) Date Survey Completed</b>  08/15/2023
<b>Name of Provider or Supplier</b>  Riverdale Park Pediatrics Pc	<b>Street Address, City, State</b>  6103 Baltimore Ave T-1, Riverdale, MD	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D6047</b>	<p><b>TECHNICAL CONSULTANT RESPONSIBILITIES</b> CFR(s): 493.1413(b)(8)(i)</p> <p>The procedures for evaluation of the competency of the staff must include, but are not limited to direct observations of routine patient test performance, including patient preparation, if applicable, specimen handling, processing and testing.</p> <p>This STANDARD is not met as evidenced by: Based on review of the competency check records for Testing Person #3, #4 and #5 along with interview, the technical consultant did not have records for 2022 showing that each was observed during routine patient test performance, including patient preparation, if applicable, specimen handling, processing and testing to check annual competency. This was confirmed during interview with the technical consultant the afternoon of the survey.</p>
<b>D6049</b>	<p><b>TECHNICAL CONSULTANT RESPONSIBILITIES</b> CFR(s): 493.1413(b)(8)(iii)</p> <p>The procedures for evaluation of the competency of the staff must include, but are not limited to review of intermediate test results or worksheets, quality control records, proficiency testing results, and preventive maintenance records.</p> <p>This STANDARD is not met as evidenced by: Based on review of the competency check records for Testing Person #3, #4 and #5 along with interview, the technical consultant did not have records for 2022 showing that each had reviews of quality control records, and preventive maintenance records to check annual competency. This was confirmed during interview with the technical consultant the afternoon of the survey.</p>

**D6050**

**TECHNICAL CONSULTANT RESPONSIBILITIES**

CFR(s): 493.1413(b)(8)(iv)

The procedures for evaluation of the competency of the staff must include, but are not limited to direct observation of performance of instrument maintenance and function checks.

This STANDARD is not met as evidenced by:

Based on review of the competency check records for Testing Person #3, #4 and #5 along with interview, the technical consultant did not have records for 2022 showing that each had direct observation of performance of instrument maintenance and function checks. to check annual competency. This was confirmed during interview with the technical consultant the afternoon of the survey.