

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 21D0211985	(X3) Date Survey Completed 04/26/2023
Name of Provider or Supplier Olney Pediatrics	Street Address, City, State 18111 Prince Philip Dr, Ste 311, Olney, MD	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2015	<p>TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(5)(6)</p> <p>(5) The laboratory must document the handling, preparation, processing, examination, and each step in the testing and reporting of results for all proficiency testing samples. The laboratory must maintain a copy of all records, including a copy of the proficiency testing program report forms used by the laboratory to record proficiency testing results including the attestation statement provided by the PT program, signed by the analyst and the laboratory director, documenting that proficiency testing samples were tested in the same manner as patient specimens, for a minimum of two years from the date of the proficiency testing event. (6) PT is required for only the test system, assay, or examination used as the primary method for patient testing during the PT event.</p> <p>This STANDARD is not met as evidenced by: Based on review of proficiency testing records and interview with laboratory staff, the laboratory did not document staff who performed the testing for each proficiency sample. Findings: 1. Six staff members signed the proficiency testing attestation statement for the third event of 2022, but the proficiency test records and laboratory records did not identify who performed testing for each individual sample; and 2. This was confirmed during interview with laboratory staff on the afternoon of the day of testing.</p>