

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 21D0219030	(X3) Date Survey Completed 02/06/2019
Name of Provider or Supplier Hillcrest Clinic Inc	Street Address, City, State 5602 Baltimore National Pike 600, Baltimore, MD	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5789	<p>TEST RECORDS CFR(s): 493.1283(b)</p> <p>Records of patient testing including, if applicable, instrument printouts, must be retained.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with laboratory staff, the lab did not keep intermediate Rh test results. Findings: 1. The laboratory documents the original patient Rh test results and the interpretation onto a handwritten log. The laboratory will then transcribe the log as a typed record; 2. The lab does not keep the original test log (intermediate handwritten test record), but discards this record and replaces it with the typed record; and 3. This was confirmed with laboratory staff on the day of survey.</p>