Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 21D0220302	(X3) Date Survey Completed 06/18/2018
Name of Provider or Supplier Dorchester County Health Department	**Street Address, City, State** 3 Cedar Street, Cambridge, MD	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6054	TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(9) The technical consultant is responsible for evaluating and documenting the performance of individuals responsible for moderate complexity testing at least annually, after the first year. This STANDARD is not met as evidenced by: Based on review of the annual competency records and interview with the testing person, the technical consultant/laboratory director did not ensure that all testing personnel received an annual competency review. Findings: 1. The laboratory currently has 2 testing personnel listed on the "Laboratory Personnel Report (CLIA)" form. 2. The annual competencies for 2018 and 2017 were reviewed. The records showed that one of the two testing personnel did not have an annual evaluation for 2018 and 2017. 3. During the survey on 06/18/18 at 12:45 PM the testing person confirmed that the annual competency had not been performed for two years for the other testing person at the laboratory.