

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 21D0893112	(X3) Date Survey Completed 07/23/2026
Name of Provider or Supplier Patient First -Lutherville	Street Address, City, State 10755 Falls Road Suite 160, Lutherville, MD	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d) (d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality. This STANDARD is not met as evidenced by: Based on record review and an interview with the technical consultant, the laboratory failed to ensure that staff did not use the chemistry level III control reagent (lot number 121183) after its March 31, 2026, expiration date. Findings: 1. The i-stat chemistry log showed that control reagent Level III lot number 121183 was in use from April 9, 2026 to April 11, 2026 (for 3 of 3 days) on the i-Stat chemistry log. 2. The finding was confirmed during interview with the technical consultant at 12:00 PM on 7/23/26.