

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 21D0922822	(X3) Date Survey Completed 06/14/2018
Name of Provider or Supplier Drs Ford & Rothman, Ctr For Cosmetic & Clinical	Street Address, City, State 18310 Montgomery Village Avenue #700, Gaithersburg, MD	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with the laboratory director, the laboratory did not have a written policy describing the system used for verifying the accuracy and reliability of interpretation of the histopathology specimens for which proficiency samples are not available. Findings: 1. The procedure did not have a written policy for evaluating the proficiency of the testing personnel at least twice a year. 2. On 06/14 /18 at 2:15 PM the laboratory director confirmed that the procedure did not have a written policy for evaluating the proficiency of the testing personnel at least twice a year. 3. The procedure should include, but is not limited to: the details of whom the samples will be exchanged with; how often samples will be exchanged; what the acceptable limits of the comparison and remedial actions to be taken when the results are not within acceptable limits.</p>