

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 21D0929322	(X3) Date Survey Completed 11/30/2020
Name of Provider or Supplier Rheumatology Associates Of Balto Llc	Street Address, City, State 1220b East Joppa Rd #310, Baltimore, MD	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on review of the federal proficiency testing data report and review of the comparative evaluation summery from American Proficiency Institute (API) proficiency testing (PT) program, the laboratory failed to successfully participate in the API PT program for chemistry testing, in which the laboratory is certified under CLIA. (Refer to D2093 and D2096)</p>
D2093	<p>ROUTINE CHEMISTRY CFR(s): 493.841(d)</p>

	<p>Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory performance and results in a score of 0 for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of the federal proficiency testing data report and review of the performance summary from American Proficiency Institute (API), the laboratory failed to return proficiency testing results to the proficiency testing program within the time frame. The following analyte was noted as failed in the 2020 2nd and 3rd event routine chemistry. Findings: 1. American Proficiency Institute 2020 2nd event routine chemistry overall score 0% 2. American Proficiency Institute 2020 3rd event routine chemistry overall score 0%</p>
D2096	<p>ROUTINE CHEMISTRY CFR(s): 493.841(f)</p> <p>Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on review of the federal proficiency testing data report and review of the performance summary from the American Proficiency Institute proficiency testing program, the laboratory failed to successfully participate in the API PT program for routine chemistry. The following analyte was noted as failed in the 2nd and 3rd event of 2020. Findings: 1. American Proficiency Institute 2020 2nd event routine chemistry 0% 2. American Proficiency Institute 2020 3rd event routine chemistry 0%</p>
D6000	<p>MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on review of the federal proficiency testing data report and review of the comparative evaluation summary from American Proficiency Institute (API) proficiency testing (PT) program, the laboratory director failed to ensure the laboratory successfully participated in the API PT program for routine chemistry testing, in which the laboratory is certified under CLIA. (Refer to D2093 and 2096)</p>
D6019	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(iv)</p> <p>The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(iv) Ensure that an approved corrective action plan is followed</p>

when any proficiency testing results are found to be unacceptable or unsatisfactory.

This STANDARD is not met as evidenced by:

Based on review of the federal proficiency testing data report and review of the comparative evaluation summary from American Proficiency Institute (API) proficiency testing (PT) program, the laboratory director failed to ensure that an approved corrective action plan was followed when the laboratory failed to successfully attain a score of 80% in the API PT program for routine chemistry testing. Findings: 1. American Proficiency Institute 2020 2nd event routine chemistry overall score 0% 2. American Proficiency Institute 2020 3rd event routine chemistry overall score 0% 3. American Proficiency Institute 2020 2nd event routine chemistry 0% 4. American Proficiency Institute 2020 3rd event routine chemistry 0%