

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 21D1032562	(X3) Date Survey Completed 01/27/2022
Name of Provider or Supplier Skin Care Specialty Physicians Llc	Street Address, City, State 1447 York Road Suite 301, Lutherville, MD	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with laboratory staff, the laboratory did not document the manufacturer's name, lot number and expiration date of test kits and stains for Mart-1, SOX-10 and CK-5 stains to ensure they are not used past expiration. Findings: 1. The laboratory did not document the manufacturer name, lot number and expiration dates for immunohistochemical stains used for patient testing; 2. This was confirmed with the histotechnologist on the afternoon of the day of survey.</p>