

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 21D1085740	(X3) Date Survey Completed 12/11/2020
Name of Provider or Supplier Ali Hendi, Md, Pc	Street Address, City, State 5454 Wisconsin Avenue Suite 725, Chevy Chase, MD	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5415	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(c)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies, as appropriate, must be labeled to indicate the following: (1) Identity and when significant, titer, strength or concentration. (2) Storage requirements. (3) Preparation and expiration dates. (4) Other pertinent information required for proper use.</p> <p>This STANDARD is not met as evidenced by: Based on reagent log review and interview with the histotech, the laboratory did not ensure that all histopathology reagents were documented on the reagent log and labeled upon opening. Findings: 1. A review of reagent logs from 06/11/2018 to 08/29/2020 showed that a bottle of "Eosin Y Stain 1%," lot # 1801907, expiration date 01/26/2020 was opened on 10/15/2019. There were no other bottles of Eosin stain documented after that date. 2. During a telephone interview on 12/11/2020 at 1:35 PM, the histotech stated that a new bottle of Eosin stain had been opened and was currently in-use and that the in-use bottle was not labeled with the date of opening. They confirmed that the new bottle of Eosin stain had not been documented on the laboratory's reagent log.</p>