Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	21D2027263	04/15/2019
Name of Provider or Supplier	**Street Address, City, State**	
Advanced Dermatology Of Maryland	2401 Research Blvd Ste 260, Rockville, MD	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6094	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(5) The laboratory director must ensure that the quality assessment programs are established and maintained to assure the quality of laboratory services provided and to identify failures in quality as they occur. This STANDARD is not met as evidenced by: Note: This is a repeat deficiency. The laboratory was cited during the re-certification survey on 4/10/2017 for not maintaining the quality assurance plan. The plan of correction stated that this would be corrected. Based on standard operating procedure manual (SOPM) and quality assurance (QA) record review and interview with the laboratory director (LD), the LD failed to follow the established QA program to assure the quality of laboratory services provided and to identify failures in quality as they occur. Findings: 1. The procedure, "Quality Assessment Review Meeting" states, "the Laboratory Director will meet quarterly with all laboratory staff to interactively discuss problems and ways to improve the quality of laboratory procedures." 2. A review of "QA Meeting" worksheets from April, 2017 to March, 2019 showed that QA meetings were held 4/10/17, 5/8/17, 1/29/18, and 1/14/19. 3. During an interview on 4/15/19 at 11:00 AM, the LD confirmed that QA meeting were not held quarterly.