

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 21D2119640	(X3) Date Survey Completed 07/25/2023
Name of Provider or Supplier Frederick P Smith, Md, Pc, Division Of Rcca	Street Address, City, State 5530 Wisconsin Avenue Suite 1640, Chevy Chase, MD	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2007	<p>TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(1)</p> <p>The samples must be examined or tested with the laboratory's regular patient workload by personnel who routinely perform the testing in the laboratory, using the laboratory's routine methods</p> <p>This STANDARD is not met as evidenced by: Based on review of proficiency test records and interview with staff, the hematology laboratory did not rotate proficiency testing among all testing staff. Findings: 1. Testing Person #2 performed the hematology proficiency testing (automated CBC) for the 1st, 2nd and 3rd events of 2022 and the 1st and 2nd events for 2023, the proficiency testing was not rotated among the other two testing persons. 2. This was confirmed during interview with staff on the afternoon of the day of survey.</p>