

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  21D2130241	<b>(X3) Date Survey Completed</b>  09/12/2023
<b>Name of Provider or Supplier</b>  Pinnacle Dermatology Lab DbA Dermatology Assoc And	<b>Street Address, City, State</b>  1110 Professional Ct Ste 201, Hagerstown, MD	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5401</b>	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on record review, observation and interview, the laboratory written procedure for manual coverslipping did not reflect actual practice in the laboratory. Findings: 1. The laboratory written procedure for manual coverslipping of slides states that the coverslipping procedure is conducted under a fume hood, but it was observed that there was no fume hood in the laboratory. 3. This was confirmed with the histotechnician during interview the morning of the survey.</p>