

|  |   |   |
|--|---|---|
| <b>Statement of Deficiencies</b>   | <b>(X1) Provider/Supplier/CLIA Identification Number</b><br><br>21D2133900                  | <b>(X3) Date Survey Completed</b><br><br>04/30/2021 |
| <b>Name of Provider or Supplier</b><br><br>Advanced Dermatology Of Maryland  | <b>Street Address, City, State</b><br><br>200 Harry S Truman Parkway Ste 400, Annapolis, MD |   |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. |   |   |

| <b>(X4) ID Prefix Tag</b> | <b>Summary Statement of Deficiencies</b>   |
|---------------------------|--|
| <b>D5217</b>              | <p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE<br/>CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by:<br/>Based on procedure manual and proficiency testing record review and interview with the regional manager, the laboratory did not ensure that proficiency testing (PT) was performed at least twice annually. Findings: 1. The procedure, "Proficiency Testing Mohs Histopatholgy" states that "Proficiency testing will be performed bi-annually. It is to be sent out and returned by end of June and Dec." 2. A review of PT records showed that no PT slides had been sent out in 2020. In an email on 4/27/2021 the regional manager stated that "with the change in lab directors and the pandemic proficiency testing was not done last year." 3. During an interview via email on 4/30/2021 at 11:30 AM , the regional manager confirmed that PT slides were not sent out in 2020.</p> |