

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 21D2133900	(X3) Date Survey Completed 02/27/2023
Name of Provider or Supplier Advanced Dermatology Of Maryland	Street Address, City, State 200 Harry S Truman Parkway Ste 400, Annapolis, MD	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6070	<p>TESTING PERSONNEL RESPONSIBILITIES CFR(s): 493.1425(b)(1)</p> <p>Each individual performing moderate complexity testing must follow the laboratory's procedures for specimen handling and processing, test analyses, reporting and maintaining records of patient test results.</p> <p>This STANDARD is not met as evidenced by: Based on review of the slide labeling procedure, slide review and interview with the supervisor, the testing personnel (TP) failed to label the patient slides with the patient's last name as required in the labeling procedure. Findings: 1. The "Analytical" section of the "Mohs Quality Assurance Manual" states: "The slide is labeled with the patient's last name, case number, stage and piece, and dated." 2. The supervisor pulled the slides for patient #FA22-008 slides for review and one of the four slides did not have the last name of the patient. The slide was labeled with "QC[quality control]" instead of the patient's last name. 3. Review of the slides from 05/15/2021 through 02/25/2023 showed that the first slide of the day of patient testing was labeled with "QC" and not the last name of the patient which is required by the "Mohs Quality Assurance Manual." 4. During the survey on 02/27/2023 at 11:30 am, the supervisor confirmed that the TP were not labeling the patient slide with the last name of the patient as required by the "Mohs Quality Assurance Manual."</p>