

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 21D2142636	(X3) Date Survey Completed 07/20/2022
Name of Provider or Supplier Anne Arundel Dermatology Berlin	Street Address, City, State 9948 Main Street, Berlin, MD	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5415	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(c)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies, as appropriate, must be labeled to indicate the following: (1) Identity and when significant, titer, strength or concentration. (2) Storage requirements. (3) Preparation and expiration dates. (4) Other pertinent information required for proper use.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with the histotechnologist, the laboratory did not ensure that the in-use histology stain reagents were labeled with the date that they were opened. Findings: 1. During a tour of the laboratory at 11:15 AM, it was observed that the opened and in-use histology stain reagents in the laboratory flammable cabinet were not labeled with the date that they were opened and put in to use. 2. During an interview on 07/20/2022 at 12:15 PM, the histotechnologist confirmed that the histology stain reagents in-use were not labeled with the date that they were opened.</p>