

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 21D2183615	(X3) Date Survey Completed 09/12/2021
Name of Provider or Supplier Aeon Technologies, Llc	Street Address, City, State 1 Technology Drive #1600, Frostburg, MD	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on observation of records and interview with the technical supervisor, the laboratory did not perform competency evaluations for the clinical consultant. Findings: 1. The personnel file for the clinical consultant did not include competency check evaluations (performed by the laboratory director) to ensure that the clinical consultant performed duties as required and in a reliable manner; and 2. This was confirmed during interview, with the technical supervisor, on the afternoon of the day of survey.</p>
D5807	<p>TEST REPORT CFR(s): 493.1291(d)</p> <p>Pertinent "reference intervals" or "normal" values, as determined by the laboratory performing the tests, must be available to the authorized person who ordered the tests and, if applicable, the individual responsible for using the test results.</p> <p>This STANDARD is not met as evidenced by: Based on review of one final test report for the reporting of a patient blood glucose result, the report did not include the reference range for the blood glucose result. Findings: 1. The glucose reference (normal) range was not reported on the one final report for Patient A who had a routine chemistry panel that included results for blood glucose. The reference ranges for all other chemistry analytes were given on the</p>

report; and 2. The missing reference range for the blood glucose was confirmed during interview with the technical supervisor during the afternoon of the day of survey.