

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 22D0682738	(X3) Date Survey Completed 03/24/2021
Name of Provider or Supplier Centre Pediatric Associates, Pc	Street Address, City, State One Brookline Place Suite 327, Brookline, MA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A CLIA recertification survey was conducted for the Centre Pediatric Associates, P.C. laboratory pursuant to the Clinical Laboratory Improvement Amendments (CLIA) of 1988 and CLIA regulations at 42 CFR 493.
D5211	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on proficiency testing (PT) review and interview with Technical Consultant (TC) on 3/24/2021, the laboratory director failed to effectively review and evaluate PT results obtained on proficiency testing performed as specified in subpart H of this part as evidenced by the following: The surveyor reviewed American Proficiency Institute (API) PT records for calendar years 2019 and 2020 (12 testing events). The review revealed that the laboratory director failed to review and evaluate 6 out of 12 final reports for the following testing events: 1. API Hematology/Coagulation 2019 Events 1, 2, and 3. 3. API Microbiology 2019 Events 1, 2, and 3. The TC interviewed on 3/24/2021 at 10:45 AM confirmed that the laboratory director failed to review and evaluate the API PT results for 2019.</p>