

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  22D0985117	<b>(X3) Date Survey Completed</b>  04/13/2023
<b>Name of Provider or Supplier</b>  Wmps At Berkshire Cosmetic & Reconstructive	<b>Street Address, City, State</b>  426 South St, Pittsfield, MA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A CLIA recertification survey was conducted for the WMPS at Berkshire Cosmetic & Reconstructive Surgery laboratory pursuant to the Clinical Laboratory Improvement Amendments (CLIA) of 1988 and CLIA regulations at 42 CFR 493. Please refer to Conditions of Participation for Clinical Laboratories 42 CFR Part 493. .
<b>D6084</b>	<p><b>LABORATORY DIRECTOR RESPONSIBILITIES</b> CFR(s): 493.1445(e)(2)</p> <p>The laboratory director must ensure that the physical plant and environmental conditions provide a safe environment in which employees are protected from physical, chemical, and biological hazards.</p> <p>This STANDARD is not met as evidenced by: Based on interview, the Laboratory Director failed to ensure that the physical plant and environmental conditions provided a safe environment in which employees were protected from physical, chemical, and biological hazards as evidenced by the following: Emergency Eyewash Station: a) On the day of the survey the maintenance records of the emergency eyewash were reviewed. The review revealed that the documentation that the eyewash was being checked was only documented for one week in January of 2023 (1/6/23) and only two weeks in February, 2023 (2/12/ and 2/27). There was no documentation that the eyewash was checked weekly in March or the first week of April of 2023. b) The Laboratory Director interviewed on 4/13/23 at 10:28 AM confirmed that documentation was not being appropriately maintained that the emergency eyewash was being routinely checked and maintained on a weekly basis .</p>