

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  22D1049113	<b>(X3) Date Survey Completed</b>  05/18/2021
<b>Name of Provider or Supplier</b>  Metromedic Walk-In Medical Center	<b>Street Address, City, State</b>  427 Plymouth Avenue, Fall River, MA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A CLIA recertification survey was conducted for the Metromedic Walk-In Medical Center laboratory pursuant to the Clinical Laboratory Improvement Amendments (CLIA) of 1988 and CLIA regulations at 42 CFR 493.
<b>D6046</b>	<p><b>TECHNICAL CONSULTANT RESPONSIBILITIES</b> CFR(s): 493.1413(b)(8)</p> <p>(b) The technical consultant is responsible for-- (b)(8) Evaluating the competency of all testing personnel and assuring that the staff maintain their competency to perform test procedures and report test results promptly, accurately and proficiently.</p> <p>This STANDARD is not met as evidenced by:  . Based on lack of record review and interview on 5/18/21, the technical consultant (TC) failed to conduct competency assessments as evidenced by the following: The surveyor asked to review competency assessments for the testing person (TP) on 5/18/21. No documentation of competency assessments for 2018, 2019, and 2020 could be provided during the time of the survey for the TP assuring they maintain their competency to perform test procedures and report test results promptly, accurately and proficiently. The TP confirmed in an interview on 5/18/21 at 10:30 AM that competency assessments were performed but not documented in 2018, 2019, and 2020. The laboratory performs approximately 5730 Routine Chemistry and Toxicology tests annually.</p>