

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 22D1098830	(X3) Date Survey Completed 04/22/2025
Name of Provider or Supplier Northeast Dermatology Associates PLLC	Street Address, City, State 138 Conant St, Beverly, MA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A CLIA recertification survey was conducted for the Northeast Dermatology Associates - Beverly laboratory on 04/22/2025 pursuant to the Clinical Laboratory Improvement Amendments (CLIA) of 1988 and CLIA regulations at 42 CFR 4
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: . Based record review and staff interview with the Histotech (HT) the laboratory failed to verify at least twice annually the accuracy of testing in the specialty of Histopathology. Findings include: 1. Record review on 4/22/2025 of the laboratory's Procedure Manual, Section 8, The Proficiency Testing and Quality Assurance Contract states "At least biannually, at least 3 cases will be randomly selected...to be evaluated for any inconsistencies." 2. Record review on 4/22/2025 of the laboratory's Quality Control binder revealed in 2023, 2024, and 2025 to date three cases for each year were all reviewed on April 1, 2025. 3. Interview with the HT on 4/22/2025 at 10:00 AM confirmed the above. The laboratory performs 1031 patient tests annually in the specialty of Histopathology</p>