

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  22D2007956	<b>(X3) Date Survey Completed</b>  02/20/2026
<b>Name of Provider or Supplier</b>  Id Of Quincy	<b>Street Address, City, State</b>  40 Industrial Park Road, Suite 202, Plymouth, MA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A CLIA recertification survey was conducted for the ID of Quincy laboratory on 02/20/2026 pursuant to the Clinical Laboratory Improvement Amendments (CLIA) of 1988 and CLIA regulations at 42 CFR 493. .
<b>D5217</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: . Based on record review and interview with the Histotechnician on 02/20/2026, the laboratory failed to verify at least twice annually procedures it performs that are not included in subpart I of this part as evidenced by the following: The surveyor reviewed the laboratory's procedure and records for the twice annual peer slide review of histopathology Mohs slide examinations for calendar years 2024 and 2025. Per the laboratory's Quality Assurance Procedure: "Proficiency testing is performed on a semi-annual basis. Cases are randomly pulled and sent to another physician for review." The review revealed that the laboratory performed the twice annual peer slide reviews on 1/30/2026 for the 1st &amp; 2nd half of 2025 - 4 cases for each half, on 4/01/2025 for the 2nd half of 2024 - 6 cases, and on 3/30/2024 for the 1st half of 2024 - 6 cases. The Histotechnician confirmed in an interview on 02/20/2026 at 10:11 A.M. that the twice annual peer slide review was not performed on a semi-annual basis for calendar years 2024 and 2025. The laboratory performs approximately 717 Mohs cases annually. .</p>