

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 22D2011968	(X3) Date Survey Completed 02/08/2023
Name of Provider or Supplier Savida Health Pc	Street Address, City, State 12 Dallaire Ave, Chicopee, MA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A CLIA recertification survey was conducted for the SAvida Health, PC laboratory pursuant to the Clinical Laboratory Improvement Amendments (CLIA) of 1988 and CLIA regulations at 42 CFR 493. Please refer to Conditions of Participation for Clinical Laboratories 42 CFR Part 493. .
D5439	<p>CALIBRATION AND CALIBRATION VERIFICATION CFR(s): 493.1255(b)</p> <p>Unless otherwise specified in this subpart, for each applicable test system the laboratory must do the following: Perform and document calibration verification procedure - (b)(1) Following the manufacturer's calibration verification instructions; (b)(2) Using the criteria verified or established by the laboratory under 493.1253(b)(3) -- (b)(2)(i) Including the number, type, and concentration of the materials, as well as acceptable limits for calibration verification; and (b)(2)(ii) Including at least a minimal (or zero) value, a mid-point value, and a maximum value near the upper limit of the range to verify the laboratory's reportable range of test results for the test system; and (b)(3) At least once every 6 months and whenever any of the following occur: (b)(3)(i) A complete change of reagents for a procedure is introduced, unless the laboratory can demonstrate that changing reagent lot numbers does not affect the range used to report patient test results, and control values are not adversely affected by reagent lot number changes. (b)(3)(ii) There is major preventive maintenance or replacement of critical parts that may influence test performance. (b)(3)(iii) Control materials reflect an unusual trend or shift, or are outside of the laboratory's acceptable limits, and other means of assessing and correcting unacceptable control values fail to identify and correct the problem. (b)(3)(iv) The laboratory's established schedule for verifying the reportable range for patient test results requires more frequent calibration verification.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to perform calibration</p>

verifications that covered the reportable range of all analytes as evidenced by the following: a) A review of urinary specific gravity calibration records for calendar year 2022 was performed. b) The review revealed that the calibration verifications performed did not cover the reportable range of 1.0010 to 1.0020. The calibration verifications performed on 4/12/22 and 10/12/22 had three (3) points. The three points were all listed as 1.0002 so did not cover the middle and upper range of the reportable range. c) The laboratory director one confirmed in an interview on 2/8/23 at 10:50 AM that the calibrations performed had not covered the reportable range of the method. d) The laboratory performs approximately 103,000 urinary specific gravity tests annually. .